

E-5

**PROCUREMENT OF GOODS
UNDER
NATIONAL SHOPPING PROCEDURES**

*(For Contracts valued less than the
equivalent of US \$30,000 each)*

**INVITATION FOR QUOTATIONS FOR SUPPLY OF
EQUIPMENT UNDER NATIONAL SHOPPING PROCEDURES**

SUB : INVITATION FOR QUOTATIONS FOR SUPPLY OF CONTINUOUS PASSIVE MOTION (CPM), MAJOR TENS, FINGER EXERCISER WITH BASE & INFRARED LAMP WITH INTENSITY CONTROL & TIMER

1. The Government of India has received a credit (Credit No –3867 IN) from the World Bank in various currencies towards the cost of the Rajasthan Health Systems Development Project and intends to apply part of the proceeds of this credit to eligible payments under the contract for which this invitation for bids is issued.
2. The Project Director, RHSDP invites sealed bids from eligible bidders for supply of Equipments listed below for RHSDP-

Item No.	Brief Description	Quantity	Delivery Period	Bid Security (in Rs.)	Warranty
G30	Continuous passive motion (CPM)	6	30 days	6400	24 months
	Major TENS	3			
	Finger exerciser with base	9			
	Infrared lamp with intensity control & timer	9			

3. **Bid Price**

- a) The contract shall be for the full quantity as described above. Corrections, if any, shall be made by crossing out, initialing, dating and re writing.
- b) All duties, taxes and other levies payable by the contractor under the contract shall be included in the total price.
- c) The rates quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- d) The Prices should be quoted in Indian Rupees only.

4. Each bidder shall submit only one quotation.

5. **Validity of Quotation**

Quotation shall remain valid for a period not less than 45 days after the deadline date specified for submission.

6. **Evaluation of Quotations**

The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which

- (a) are properly signed ; and
- (b) confer to the terms and conditions, and specifications.
- (c) Past performance
- (d) EMD in form of D.D./ Bankers Cheques in favour of PD, RHSDP payable at Jaipur as per para 2
- (e) The Quotations would be evaluated together for package G30.

7. **Award of contract**

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price. The successful bidder shall furnish the performance security @ 5% of the contract value in form of DD/ Banker's cheque or Bank guarantee and shall remain valid up to after 60 days completion of warranty period.

7.1 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.

7.2 The bidder whose bid is accepted will be notified of the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order.

8. Payment shall be made immediately after delivery of the goods.

9. Normal commercial warranty/ guarantee shall be applicable to the supplied goods.

FORMAT OF QUOTATION *

Sl. No.	Description Goods	Qty.	Quoted Unit Rate in Rs.	Total Amount		VAT/ Sales tax*
				In Figures	In Words	
1	Continuous passive motion (CPM)	6				
2	Major TENS	3				
3	Finger exerciser with base	9				
4	Infrared lamp with intensity control & timer	9				

Note: VAT/ sales tax should be quoted separately. Evaluation will be done on Unit price.

Total Amount : Rs.

We agree to supply the above goods in accordance with the technical specifications for a total contract price of Rs.(amount in figures) (Rs. amount in words) within the period specified in the Invitation for Quotations.

We also confirm that the normal commercial warrantee/guarantee of months shall apply to the offered goods.

Signature of Supplier

CONTINUOUS PASSIVE MOTION (LOWER EXTRIMITY)(CPM)

TECHNICAL SPECIFICATIONS:

It Must Have Timer (digital) 0 – 30 Minute

It Must Have range of motion 0 – 120⁰

It Must Have Provided along with Patient Safety Switch

It Must Have delay time 0-10 sec for both extension & Flexion

Low noise level for smooth operations.

Audio visual alarm.

Technical Specifications of DUAL Channel T.E.N.S. COMBINED WITH M.S (SD CURVE)

- It must have with following such as Faradic, Surge Faradic, Galvanic, Int Galvanic, Cont TENS & Surge TENS
- It must have 2 Independently controlled output Channel FOR T.E.N.S OR M.S
- It must have maximum amplitude of 99 mA
- It must have phase duration from 0.01 to 300 msec In Ten Steps
- It must have frequency from 2-250 Hz
- It must have Pulse Width 100 Micro Sec to 800 Micro Sec
- It must have timer from 1 to 60 Minutes
- It must have LCD Display
- It must have program compatibility of one program per output mode

- It must have power off, Adjustable warning volume
- It should be supplied with 4 color coated leads, 4 Rectangular Electrodes 4 Sponges mains lead, and operating manual

NON LUMINOUS INFRA RED LAMP 150 W (TABLE TOP)

- It should have Good Heat Penetration
- It should be a Non- luminous and table top model
- It should have protective grid in front part of the lamp
- It should have handle on the top of the lamp
- It should have manual treatment timer of 0-60 minute
- It should be of 150W-400W
- It should have wavelength of 1-5 microns

Finger Exerciser with Base

Finger Exerciser with Wooden frame and SS spring and SS frame.

**Consignee List for CPM (Continuous Passive Motion), Major Tense,
Finger exerciser with base & Infrared lamp with intensity control & timer**

Sr. No.	Name of City/Town	CPM (Continuous Passive Motion)	Major Tense	Finger exerciser with base	Infrared lamp with intensity control & timer
1	GANGANAGAR	2	1	3	3
2	JAISALMER	2	1	3	3
3	SIKAR	2	1	3	3
	Total	6	3	9	9